

27 June 2023

Dr. Dr/Za-hu

TEST CERTIFICATE

1211 P 23

for Messrs: Neenah Coldenhove B.V.
D.W. Van Vreeswijklaan 9
6961 LG Eerbeek / The Netherlands

Product: Sterilization paper for use in the manufacture of paper bags and in
the manufacture of pouches and reels

Designation: Medicol Steam 68 gsm

The product manufactured by the above-mentioned company is intended to be used as packaging material for terminally sterilized medical devices which are to be sterilized by steam (134 °C), by ethylene oxide or by gamma irradiation (50 kGy).

The sample material submitted for analyses was examined by us according to

ISO 11607-1:2019-02 section 5.2.3, analysis according to DIN 58953-6:2016-12 and
DIN EN 868-3:2017-05 section 4.2.

The results are summarized in the test report no. 12311/40 of 27 June 2023 issued by ISEGA Forschungs- und Untersuchungsgesellschaft mbH Aschaffenburg, Germany.

The product "Medicol Steam 68 gsm" meets the requirements for packaging for terminally sterilized medical devices according to

ISO 11607-1:2019-02 section 5.1.6 a) microbial barrier,
section 5.1.6 c) physical and chemical properties,
section 5.1.6 e) compatibility with respect to the intended
sterilization processes

with regard to the above-mentioned sterilization processes and specification.



(Zahn)
Head of
Physical Material Testing