

Dr. Ralph Derra

Öffentlich bestellter und vereidigter Sachverständiger für
Verpackungsmaterialien, Boden- und Luftanalysen;
Sachverständiger in der Wasseranalytik

19 December 2023

Dr. Dr/Za-pa

TEST CERTIFICATE

1229 P 23

for Messrs: Neenah Coldenhove B.V.
D.W. Van Vreeswijklaan 9
6961 LG Eerbeek / The Netherlands

Product: Paper for low temperature sterilization processes

Designation: Medicol ETO 60 gsm

The product manufactured by the above-mentioned company is intended to be used as packaging material for terminally sterilized medical devices which are to be sterilized by ethylene oxide or by gamma irradiation (50 kGy).

The sample material submitted for analyses was examined by us according to

ISO 11607-1:2019-02 section 5.2.3, analysis according to DIN 58953-6:2016-12 and
DIN EN 868-6:2017-05 section 4.2.

The results are summarized in the test report no. 12311/43 of 19 December 2023 issued by ISEGA Forschungs- und Untersuchungsgesellschaft mbH Aschaffenburg, Germany.

The product "Medicol Steam 60 gsm" meets the requirements for packaging for terminally sterilized medical devices according to

ISO 11607-1:2019-02 section 5.1.6 a) microbial barrier,
section 5.1.6 c) physical and chemical properties,
section 5.1.6 e) compatibility with respect to the intended
sterilization processes

with regard to the above-mentioned sterilization processes and specification.



P. Zahn

(Zahn)

Head of
Physical Material Testing